

ECT Cutter

Type „Billerud“

Code: N.312.025

Usage

To prepare samples for the ECT (Edge Crush Test).

Applicable standards

- ISO 3037

Device description

This highly precise “Billerud” Cutter is equipped with two parallel guided razor blades, which cut out very clean and parallel samples of 25 mm in width and 100 mm length from corrugated board. When the sample to be cut is laid onto the safety guide and the lever is actuated, the blades are moved via compressed air and cut out the sample.

Specifications

- highly precise razor blades
- parallel and clean cutting
- sample width: 25 mm
- sample length: 100 mm
- up to 7 mm sample height
- mechanical timer
- safety guides
- acrylic glass safety cover
- pneumatic cutting
- CE certificate

Connections

- Air: Ø6 mm (max. 400 kPa)

Delivery content

- ECT Cutter Type “Billerud”
- support plate
- Ø6 mm air hose
- user manual
- CE label

